

EC Certificate



Production Quality Assurance Directive 93/42/EEC on Medical Devices, Annex V

Registration No.: DD 1023668-1

Manufacturer: Kriomedpol Sp. z o.o.
Ul. Warszawska 272
05-082 Stare Babice
Poland

Products: - Cryotherapy devices for local cryostimulation
- Cryosurgery - tissue cryodestruction devices

Replaces EC Certificate number DD 60119069 0001

TÜVRheinland

The Notified Body hereby declares that the requirements of Annex V of the directive 93/42/EEC have been met for the listed products. The above named manufacturer has established and applies a quality assurance system, which is subject to periodic surveillance, defined by Annex V, section 4 of the aforementioned directive. For placing on the market of class IIb and class III devices covered by this certificate an EC type-examination certificate according to Annex III is required.

Report No.: 84949402-20

Effective date: 2021-05-25

Expiry date: 2024-05-26

Issue date: 2021-05-25

J. Pyclik



Jarosław Pyclik
TÜV Rheinland LGA Products GmbH
Tillystraße 2 · 90431 Nürnberg · Germany

TÜV Rheinland LGA Products GmbH is a Notified Body according to Directive 93/42/EEC concerning medical devices with the identification number 0197.